

SYMBOLS

 General warning sign

 Manufacturer

 Catalog number

 Separate collection required (WEEE)

 Observe instructions for use

 Date of manufacture

 Serial number

 0197 European Conformity mark

 Note

 Importer

 Batch code

 Suitable for thermal disinfection

 Autoclavable at 134°C

 Authorized representative in the European Community

INTENDED PURPOSE

The handpieces in combination with a drive system and corresponding rotating instruments are used for milling, grinding and drilling in the ENT field. The intended use is obvious to the trained user.

CONTRA INDICATIONS

Relative or absolute contra indications can arise from the general medical diagnose, or in special cases by a significantly increased risk to the patient through the use of motor-driven devices. Relevant cases in the literature must be taken into consideration.

INTENDED USERS

Intended users are trained and qualified personnel, in professional settings (e.g. hospital, ambulatory).

AMBIENT CONDITIONS	TRANSPORT AND STORAGE	DURING USE
Relative humidity	max. 90%	max. 80%
Temperature	0 °C – 50 °C	10 °C – 30 °C
Atmospheric pressure	700 hPa – 1’060 hPa	800 hPa – 1’060 hPa

SAFETY INFORMATION



The handpiece and accessories are supplied by us non-sterile. The handpiece and accessories must be cleaned, disinfected and sterilised before first use and immediately after each use!

Prior to using the product, before startup, and before operation, the user must always ensure that the product and accessories are in good working order and are clean, sterile and operational.

Improper use or repair of the product, or failure to observe these instructions, relieves NOUVAG from any obligation arising from warranty provisions or other claims.



The use of the product other than that for which it was designed is not permitted. The responsibility is solely carried by the operator.

Manipulate the handpiece only when the motor is at a standstill.


Never operate the clamping mechanism during operation or without a clamped tool or spacer pin.

The device shall only be operated by qualified and trained personnel.

Do not clean the handpiece with compressed air!

The handpiece may be operated up to a maximum of 80’000 rpm.

OVERVIEW



USE

ATTACHING THE COOLING TUBING SET



Attach the cooling clip to the handpiece.



Connect the irrigation tube with the cooling clip.



Attach the tube fixing clip (REF 1881) to the irrigation tube and clip it to the handpiece.



INSERTING AND EXCHANGING THE TOOL



Hold handpiece with one hand at the lower part and open the clamp by turning the tightening sleeve with the other hand with a counter motion until it clicks.




Extricate spacer pin or previously used tool



Insert new tool.



Lock tightening sleeve by twisting in the opposite direction.



Check seating of the tool by gently pulling it.

REPROCESSING INSTRUCTIONS



In relation to patients with Creutzfeldt Jakob disease or its variant (vCJK) no responsibility can be assumed for re-use of the handpiece. The Robert-Koch Institute recommends removing used products from circulation after use in order to avoid infecting other patients, users and third parties.



Never clean the handpiece in an ultrasonic bath! This impairs the functionality of the handpiece.

Reprocessing restrictions	Frequent reprocessing has only a limited impact on the handpiece. The end of the products service life is normally determined by wear and damage through use. The handpiece is designed for 250 sterilization cycles.
General handling	1. The handpiece and accessories must be thoroughly cleaned, disinfected and sterilised before initial operation (products straight from the factory) and also immediately following each use. Only a cleaned and disinfected handpiece enables correct sterilisation! 2. The handpiece and accessories should always be treated with utmost care when being transported, cleaned, serviced, sterilised and stored. 3. We recommend the use of mildly alkaline and enzymatic cleaners with as low a silicate content as possible to avoid staining (sili-catization) on the handpiece and accessories. 4. Only commercial grade DGHM-/VAH-listed agents may be used for cleaning and disinfection. See these agent manufacturers' specifications for the method of use, action time and suitability of disinfection and cleaning substances. 5. Operating instructions for the equipment and chemicals etc. used during reprocessing must be strictly followed. 6. Dosage of chemicals, exposure times and exposure temperatures for cleaning and disinfection must be strictly followed. 7. The end of product life may be reached before the 250 sterilization cycles in case of excessive wear and damage from use. 8. Do not overload the washer. Avoid rinsing blind spots. Ensure secure storage in the machine. 9. Observe the regulations valid in your country for the reprocessing of medical devices. 10. NOUVAG recommends the use of a screen basket with rinsing bar from 3mach (NOUVAG REF 51401), a reusable container for convenient preparation and storage (including transport) of the products. The screen basket can be used for safe storage of the products during the rinsing process as well as during and after sterilization until the products are used. The screen basket is suitable for use with sterilization paper or a rigid sterilization container. It does not have a barrier effect on its own to protect sterility.
Preparation at the point of use	After surgery immediately remove blood, secretion, tissue and bone residue with a disposable cloth/paper towel, do not allow to dry! Dried residues cause corrosion.
Safe-keeping and transport	Contaminated products must be stored and transported to the preparation site in a closed container to avoid damaging the products and the contamination of the environment.
Cleaning and disinfection, pre-cleaning	Remove drill bits, cooling tubes, tube retaining clips and cooling clip. Wash off any visible soiling with water. 1. Wipe the handpiece and accessories with a damp disposable cloth/paper towel, removing all visible dirt. 2. Brush the handpiece and accessories under running tap water using a soft brush (e.g. Insitumend GmbH, REF MED100.33). 3. Rinse the outer surface of the handpiece and accessories for 10 seconds with a water pressure gun at a pressure of at least 2.0 bar (e.g. HEGA Medical, REF 6010 or REF 7060). Local tap water is sufficient for this purpose, since the last step is always a machine cleaning with deionized water, so possibly hard water with lime traces from the pre-cleaning cannot remain on the handpiece. 4. Flush the cooling clip with a cleaning gun with jet nozzle attachment (e.g. REF 4312) for at least 30 seconds.

Cleaning	Mechanical cleaning 1. After pre-cleaning, place the handpiece on a suitable attachment. Place the small parts (bur, tube holder clips and cooling clip) in the strainer basket. 2. Mechanical cleaning is only successful if the pre-cleaning, described above, is adhered to! 3. Cleaning is done using the Vario TD programme in the cleaning and disinfection unit (CDU). For the cleaning process it is advisable to use DI water (fully desalinated water). 4. After completing the cleaning program (incl. thermal disinfection) check the hand-piece and accessories for visible contamination in the grooves and gaps. Repeat clean-ing if necessary.	Automatic cleaning process (Vario TD programme) 1. Pre-clean with cold water for 4 minutes. 2. Empty 3. Clean for 5 minutes at 55°C with 0.5% alkaline or at 40°C with 0.5% enzymatic cleaner. 4. Empty 5. Neutralise with cold water for 3 minutes. 6. Empty 7. Inter-rinse for 2 minutes with cold water. 8. Empty
Disinfection	Mechanical disinfection The cleaning/disinfection unit has a thermal disinfection programme which follows after the cleaning. When performing mechanical thermal disinfection, give due consideration to the national requirements relating to the A0 value (see DIN EN ISO 15883-1). We recom-mend an A0 value of 3.000 for the handpiece and accessories. Disinfection must be carried out with DI water.	 Warning When inadequately rinsed or exposed to the disinfec-tant or detergent for too long, the handpiece and acces-sories can corrode. Please see the corresponding deter-gent and disinfectant’s package insert for dwell times.
Drying	Mechanical drying Drying of the handpiece and accessories through the drying cycle of the cleaning/disinfec-tion unit’s (CDU). If required, manual drying can also be achieved by using a lint-free cloth. Pay particular attention to the grooves and spaces between the handpiece and accessories. Then spray the handpiece and accessories again with lubricant. Every CDU must provide a corresponding drying procedure through the manufacturer (see ISO 15883-1). Please follow the corresponding CDU-manufacturer’s directions and oper-ating instructions.	Manual drying Set up the handpiece vertically, separate from the cool-ing clip, so that water can run out more easily. Allow the handpiece and accessories to dry for at least 30 minutes. Then spray the handpiece and accessories again with lubricant.
Manual cleaning	1. After pre-cleaning, place the handpiece in an immersion bath with enzymatic cleaner for 15 minutes. Clean the accessories in an ultrasonic bath for 15 minutes. Follow the instructions of the cleaning agent manu-facturer. 2. Clean products completely with a soft brush under running drinking water. Rinse cavities and lumens, if any, intensively (>30 sec.) with a water pressure gun (or similar). 3. To remove the cleaning agent, rinse products under running tap water (drinking quality) (>30 sec.).	 Warning Do not clean the handpiece in an ultrasonic bath!
Manual disinfection	After cleaning, immerse the products in a bath of suitable disinfectant for 5 minutes. Make sure that all surfaces are covered with the disinfectant. Follow the instructions of the disinfectant manufacturer. After disinfection, rinse all products thoroughly with deionised water (>1 min.) to remove the disinfectant.	
Inspection and care	Carry out a visual inspection for damage, corrosion and wear. After cleaning and disinfecting, spray the handpiece with lubricant and wipe with a lint-free cloth moistened with DI water (see instructions on spray can). Check the cooling tubes for blockages and repeat the cleaning cycle if necessary. Then refit the cooling clip to the handpiece.	
Sterilisation	Sterilisation of the products is performed with a fractionated pre-vacuum steam sterilisation process (in accordance with DIN EN 13060 / DIN EN 285) giving due consideration to the respective national requirements. Minimum requirements: 1. Pre-vacuum phases: 3 2. Sterilisation temperature: minimum 132°C – maximum 137°C (within the sterile band) 3. Holding time: At least 5 minutes (full cycle) 4. Drying time: At least 10 minutes When sterilising several products during one sterilisation cycle, do not exceed the maximum steriliser load (see manufacturer’s details). A drying cycle must be added in the case of autoclaves without a vacuum function. After sterilisation an immaculate sterilisation result must be detected by examin-ing the appropriate indications. According to the Robert-Koch Institute preparation ends with the documented release for use of the medical device. If the sterilised handpiece is not used immediately after sterilisation, it must be labelled with the sterilisation date on the packaging.	
Storage	Storing the sterile packaging The sterilised product must be stored away from dust, humidity and contamination. During storage, direct sunlight should be safely avoided. After the expiry date has passed, do not use the product any longer.	Handling the sterile packaging Before taking out the product, check for the packaging to be intact. When taking out the product, follow the respective aseptic procedures.
Information for validating the preparation	The above preparation process has been verified by a validated procedure. The following materials and machines were used: 1. Alkaline cleaner: Neodisher® Mediclean; Chemische Fabrik Dr. Weigert GmbH & Co. KG 2. Enzymatic cleaner: Neodisher® MediZyme; Chemische Fabrik Dr. Weigert GmbH & Co. KG 3. Cleaning and disinfection unit: Miele G 7836 CD 4. Rack trolley: Miele E429 5. Strainer basket/flush socket bar: 3mach (NOUVAG REF 51401) 6. Autoclave: Selectomat 666-HP (MMM) 7. Sterile packaging: Sterisheet 100; Broemeda Amcor Flexibles GmbH Chemicals and machines other than those mentioned can also be used. In such a case consult the manufacturers or suppliers to find out whether their products confer the same performance as the products that the procedure was validated with. If you should opt for a different procedure for repro-cessing to the one given above, you are required to correspondingly establish the suitability.	



There is no experience available from conducting other sterilisation procedures such as plasma sterilisation, low temperature sterilisation procedure, etc. Users bear full responsibility if they use a procedure which differs from the validated sterilisation procedure described!



Please also comply with the applicable legislation in your country and the medical practice or hospital’s hygiene rules. This especially applies to the varying requirements for an effective inactivation of prions.

MALFUNCTIONS AND TROUBLESHOOTING


PROBLEM	CAUSE	SOLUTION
Motor runs, but tool does not move.	Handpiece not opti-mally coupled to motor.	Press the handpiece onto the motor until it snaps into place. Check the seat with a counter movement.
Tool does not run smoothly.	Tool not optimally clamped.	Open the chuck, fix tool properly and close chuck.
Handpiece is loud.	Ball bearings not oiled or dirty.	Spray the handpiece with lubricant.

INFORMATION ON DISPOSAL



When disposing of the device, device components and accessories, the regulations issued by the legislator must be followed. Used electrical and electronic equipment is hazardous waste and must not be disposed of with household waste.

TECHNICAL DATA

REF	1903nou	1904nou	1905nou	1906nou	1907nou	1908nou
						
Shape	straight	straight	straight	angled	angled	angled
Transmission ratio	1:1	1:1	1:1	1:1	1:1	1:1
Torque max.	3 Ncm	3 Ncm	3 Ncm	3 Ncm	3 Ncm	3 Ncm
Speed max.	80'000 rpm	80'000 rpm	80'000 rpm	80'000 rpm	80'000 rpm	80'000 rpm
Coupling	ISO 3964	ISO 3964	ISO 3964	ISO 3964	ISO 3964	ISO 3964
Shank-Ø	2.35 mm Type 1, ISO 1797	2.35 mm Type 1, ISO 1797	2.35 mm Type 1, ISO 1797	2.35 mm Type 1, ISO 1797	2.35 mm Type 1, ISO 1797	2.35 mm Type 1, ISO 1797
Appropriate tool length	70 mm	95 mm	125 mm	70 mm	95 mm	125 mm
Weight, incl. cooling clip	85g	85g	90g	110g	115g	120g

POST MARKET SURVEILLANCE



If you have any complaints in relation to the use of the medical device, please contact the manufacturer immediately by e-mail complaint@nouvag.com or by phone.
In order to provide adequate information, please complete the complaint form:
<https://nouvag.com/en/service/complaint-form/>

MANUFACTURER AND SERVICE POINTS



Switzerland
NOUVAG AG
St. Gallerstrasse 25
9403 Goldach



Germany
NOUVAG GmbH
Schulthaissstrasse 15
78462 Konstanz



0197

Phone +41 71 846 66 00
info@nouvag.com
nouvag.com

Phone +49 7531 1290-0
info-de@nouvag.com
nouvag.com